

Posey® Hand Control Mitts Application Instructions



Rx ONLY

REF 2810 Soft Hand Mitts


REF 2820 Rigid Hand Control Mitts - S,M,L sizes available

DESCRIPTION OF PRODUCT: A full pocket mitt without finger separators.

Application Instructions:

(Repeat Steps 1-6 For Each Mitt):

1. Insert the patient's hand into the mitt, palm down.
2. Wrap the wrist strap around the smallest part of the patient's wrist.
3. Where available, thread the strap through the anchoring loop (Fig. 1).
4. Thread the strap through the plastic ring.
5. Secure with the hook and loop fastener.
6. Slide ONE finger (flat) between the device and the inside of the patient's wrist to ensure proper fit. The strap must be snug, but not compromise circulation.



Indications for Use:

- Patients assessed to be at risk of disrupting life-saving treatments (e.g., chronic tube pulling).
- Patients assessed to be at risk of line pulling, which may prevent monitoring of vital signs.
- Patients whose picking, pulling, scratching, or peeling exacerbates a skin condition, causes self-injury, or compromises wound site integrity.

Contraindications:

- **DO NOT** use on a patient who is or becomes highly aggressive, combative, agitated, or suicidal.
- **NEVER** use mitts on a patient:
 - If an IV or wound site could be compromised by the device; or
 - With a dislocation or fracture on the affected limb.See the Posey Catalog for other options for such a patient.

Adverse Reactions:

- Severe emotional, psychological, or physical problems may occur if: the applied device is uncomfortable, or it severely limits movement. If symptoms of these problems ever appear for any reason, get help from a qualified medical authority and find a less restrictive product or intervention.

WARNING

ADDITIONAL OR DIFFERENT BODY OR LIMB RESTRAINTS MAY BE NEEDED (See Posey Catalog):


- To reduce the risk of the patient getting access to the line/wound/tube site;
- To prevent the patient from flailing or bucking up and down and causing self-injury.

MONITOR PER FACILITY POLICY. Check to ensure that:

- Mitts are properly secured. If applied too tightly, circulation will be restricted; if applied too loosely, the patient may be able to slip his or her limb from the device;
- Mitts are intact, not torn or damaged, and hook and loop closes securely.
- The patient cannot use his or her teeth or otherwise remove the device and inflict self-injury;
- Monitor closely when the patient is out of bed. Patients who ambulate while wearing this device may be at risk of injury from a fall.
- **DO NOT** allow patients to ingest mitt material; if beads are ingested by patient he/she may experience temporary discomfort. Beads are non-toxic.

BED SAFETY:

- **ALWAYS** use Hospital Bed Safety Workgroup (HBSW) (<http://www.fda.gov> search keyword "HBSW") compliant side rails in the UP position and fill ALL gaps to reduce the risk of entrapment.
- Use side rail covers and gap protectors to help prevent the patient's body from going under, around, through or between the side rails. A failure to do so may result in serious injury or death if a patient becomes suspended or entrapped. Posey offers a full range of side rail pads and/or gap protectors to cover gaps.



ADDITIONAL SAFETY AND LAUNDERING
INSTRUCTIONS ON OTHER SIDE



Posey Company • 5635 Peck Road, Arcadia, CA 91006-0020 USA
Phone: 1.800.447.6739 • 1.626.443.3143 • Fax: 1.800.767.3933 • www.posey.com



MDSS GmbH
Schiffgraben 41
D-30175 Hannover, Germany



⚠ WARNING: ALWAYS Monitor patients per facility policy. Improper application or use of any restraint may result in serious injury or death.

RX ONLY. Federal law (USA) restricts this device to sale by or on order of a physician. For use in a licensed healthcare facility only.

STAFF TRAINING: Posey offers training aids at no charge. Contact Posey online at www.posey.com or call toll-free at 1.800.447.6739.


BEFORE APPLYING ANY RESTRAINT:

Follow your hospital's restraint policies and procedures which are in compliance with CMS guidelines and state laws.


⚠ ADDITIONAL WARNINGS:

1. **ALWAYS monitor patient per facility policy. Be aware that constant monitoring may be required for:**

- Aggressive or agitated patients; and
- Patients in a supine position; ensure that being in this position would not cause any complications for those who are at risk of vomiting while being restrained.



2. **NEVER alter or repair this product. ALWAYS inspect before each use:** Check for broken stitches or parts; torn, cut or frayed material; or locks, buckles, or hook-and-loop fasteners that do not hold securely. DO NOT use soiled or damaged products. Doing so may result in serious injury or death. Dispose of damaged products per facility policy for BIOHAZARDOUS material.



3. **ALWAYS secure straps to a part of the frame that moves with the patient (not foot/head board or siderails, out of the patients reach, using quick-release ties (see drawing below) or buckles.** These allow easy release in the event of an accident or fire. Test to make sure straps cannot tighten, loosen, or slip and create excess slack. If this occurs, the patient may slide off the chair or bed, increasing the risk of serious injury or suffocation. Restraint release is an important part of facility fire and disaster drills. Straps can be cut with scissors in an emergency.

4. **NEVER secure restraint strap to side rail or head/foot board.**

5. **NEVER use Posey products on toilets, or on any chair or furniture that does not allow proper application as directed in the Application Instructions. DO NOT use at home.**



6. **NEVER expose this product to open flame, fire, smoking materials, or high heat sources.** Some products may melt or ignite and burn. The facility smoking/no smoking policy should be strictly enforced.

7. **NEVER use a Posey product as a seat belt in a moving vehicle.**

Posey products are not designed to withstand the force of a crash or sudden stop.



8. Test Zippers or hook-and-loop fasteners before each use. DISCARD device if it does not fasten securely.

LAUNDERING INSTRUCTIONS (if applicable):

- Fasten all buckles and locks to reduce risk of damage during wash and dry cycles. DO NOT put buckles or locks through extractors. For maximum life, launder in a laundry bag.
- Before laundering, zip up and turn the product inside out to protect zipper.
- Hook-and-loop fasteners may collect lint after repeated use or laundering, reducing grip strength. Fasten the "hook" to the "loop" before laundering to help prevent lint buildup. As needed, use a stiff-bristle brush to remove lint from the "hook" side.
- These products, other than foam products, can be machine washed under CDC* guidelines for material soiled with blood or bodily fluid.
- For non-contaminated material, use lower temperature wash and dry cycles to extend product life.
- For foam products:



BIOETHANE:


- Sanitize by submerging the entire product in 70% isopropyl alcohol for 10 minutes; or
- Apply an OSHA approved intermediate level disinfectant per manufacturer instructions. After cleaning, products MUST be rinsed with water to remove any residual chemicals.
- Make sure products are completely dry before use.
- Stainless Steel Locks: Posey recommends that locks be treated with a dry-film lubricant (silicone) after each cleaning. For best results, spray a small amount into the lock. Work the action of the lock with a key several times to ensure that all the surfaces are well lubricated.

STORAGE AND HANDLING:

- This device is designed for use in normal indoor environments.
- This device may be stored in ambient warehouse temperatures at normal humidity levels. Avoid excess moisture or high humidity that may damage product materials.

*www.cdc.gov

HOW TO TIE THE POSEY QUICK-RELEASE TIE



1. Wrap the strap once around a movable part of the bed frame leaving at least an 8" (20 cm) tail. Fold the loose end in half to create a loop and cross it over the other end.
2. Insert the folded strap where the straps cross over each other, as if tying a shoelace. Pull on the loop to tighten.
3. Fold the loose end in half to create a second loop.
4. Insert the second loop into the first loop.
5. Pull on the loop to tighten. Test to make sure strap is secure and will not slide in any direction.
6. Repeat on other side. Practice quick-release ties to ensure the knot releases with one pull on the loose end of the strap.