

## Characteristics

Powder-free synthetic contains no natural rubber latex proteins. Soft, flexible, PVC polymer is bonded with a polyurethane interior, creating a wonderfully comfortable non-latex glove.

*DermAssist vinyl does not contain phthalate plasticizers, nor any materials included on the “California Proposition 65” list.*



## Non-Sterile Exam Glove

**DermAssist®**

**Vinyl  
Series 161**



## PRODUCT DETAILS

SIZE	ITEM NO.	PACKAGING	DESCRIPTION
XS	161050	100 Gloves/box, 10 boxes/case	Gloves, Exam, Vinyl, Non Sterile, Powder-Free
S	161100	100 Gloves/box, 10 boxes/case	
M	161200	100 Gloves/box, 10 boxes/case	
L	161300	100 Gloves/box, 10 boxes/case	
XL	161350	100 Gloves/box, 10 boxes/case	

## Product Attributes

- Enhanced Formulation
- Non-Latex

## Benefits

- Improved Wet/Dry Grip
- No Risk of Latex Allergens

*Product Solutions You Trust*

## *DermAssist® Synthetic Powder-Free Vinyl Exam Gloves*



- 100 Qty (By Weight)
- Single Use
- Ambidextrous
- Non-Sterile

DermAssist® Vinyl is manufactured in compliance with multiple international standards, including the following:

Designation	Standard
ASTM D5250	Standard Specification for Poly(vinyl chloride) Gloves for Medical Application
ASTM D5151	Standard Test Method for Detection of Holes in Medical Gloves
ASTM F1671	Standard Test Method for Resistance of Materials Used in Protective Clothing to Penetration by Blood-Borne Pathogens

Average Length	Average Palm Thickness	Average Finger Thickness
9.5 in ♦ 240 mm	4.0 mil ♦ 0.10 mm	5.1 mil ♦ 0.13 mm

Tensile Strength & Elongation	Before Aging	After Accelerated Aging
Tensile Strength (Mpa)	18	17
ASTM Requirement Min. (Mpa)	11	11
Elongation (%)	410	400
ASTM Requirement Min. (%)	300	300



Innovative Healthcare Corporation  
is certified to ISO 13485:2003 QMS  
for medical devices.